


What else is needed for a detector test with the temporary lock?



- Lock ready
- Dewar ready
- Infrastructure for mounting dewar to lock available
- Detectors ready and available
- Cabling on detector string available
- FE electronics ready and available

Lock ready:

Available by this year

Dewar ready:

Flange and valves need to be designed and produced →  
MPI-K

Infrastructure for mounting dewar to lock available

Temporary lock has to be jacked up by 0.5m → MPP

- Detectors ready and available

There is prototype detectors. How many? Holders?

→MPI-K

Prototype from LNGS → Glove box

MPP → Roland → Holder

- Cabling on detector string available

Matrix available. Will be critical test!

FPC not likely to be changed to coax cables soldered pogo pins

- FE electronics ready and available

3 fold PZO available? Integration to string → LNGS, Milano

- DAQ, etc.

HV, FADC (10ms), LV by Carla


- Working group:


MPI-K: Dewar, cryo infrastructure

MPP: Lock System incl. complete cable tree, jacking up lock

LNGS, Milano: Front end electronics Preamps, Filter.

Manpower: LNGS? MPI-K? MPP? Others?





Floor

Lower edge of support structure