





Installation of Cryostat's Thermal Insulation

K.T.Knöpfle
MPI Kernphysik, Heidelberg
ktkno@mpi-hd.mpg.de

GERDA Collaboration Meeting at Isola / LNGS
15 - 16 June 2009



actual realization



Extruded polystyrene foam (XPS):
Jackodur 300 FTD 2630x600x6 mm
(~0.035-0.04 W/(mK) : factor ~8 better thermal isolation than Lexan)
(long term water absorption <1 vol%)

- ▶ to be attached with (double-sided) adhesive tape
DUPLOCOLL 268

Jackodur / VM2000 test report by I. Luca & N. Stefano

ICP-MS measurements (Ca/K/Na 14/5/4ppb)

suspended particle counting


water absorption (<5vol%), pH (~5.8-5.9), conductivity

► all results acceptable



vertical wall - two options for VM2000 cover:


1)




better geometry
difficult to install

2)

worse geometry
easy to install



edges of VM2000 to be sealed with Scotch-Weld DP100 epoxy



s3

s2

s1

segments to be attached
with adhesive tape and
eccentrics; latter screwed to
metal rings.


bottom to be planked with
30 degree segments, i.e.
12 pcs in sector s1 - s3 .

Isola, 16 June 2009

Cry



The three segment types to be cut out of the Jackodur panels – templates avail.



Conclusions:

Jackodur available

Templates, tape, epoxy, eccentrics available

VM2000 foil available (Tübingen)

Tasks and schedule (not time critical):

test installation Jackodur at vertical wall

this week

test installation of J. at bottom

this week

install Jackodur at bottom

week 27 ?

cover complete cryostat surface
with VM2000

together with muon
final works